

### HPNCBES

- Call Origination Indicator Lamp
- Call Reset from Origination Point Only
- High Visibility Large Blue Button
- Heavy Duty Switch Contacts
- Faceplate Rated 94V-0 by UL
- Non-Conductive Mounting Screws

### DESCRIPTION

Wall recessed or surface mounted blue combination call and reset button with indicator lamp to be used by hospital staff members to summon help in emergencies (primarily in Cardiac Arrest situations).

### METALCAL (AVAILABLE SEPARATELY)

A self-adhesive "STAFF USE ONLY" metalcal, Part No. A44189 (shown in photo), is available to be affixed to the faceplate where required.

### CONTROLS AND INDICATORS

Integrated call and reset rocker type blue button with replaceable indicator lamp.

### SWITCH

Positive action toggle type with heavy duty independent contacts.

### CODE BLUE FUTURA II



### INSTALLATION

Mounting bracket will permit correction of minor misalignment of backbox in wall.

### HOUSING AND FINISH

**Faceplate:** high impact, molded thermoplastic faceplate finished in off-white supplied with nylon screws for mounting. **S41 Housing:** metal surface housing finished in alumex. **W41 Backbox:** metal backbox supplied by others.

### DIMENSIONS

**Faceplate:** 2-3/4" (6.99cm) wide x 4-1/2" (11.43cm) high; **S41 Housing:** 2-7/8" (7.30cm) wide x 4-5/8" (11.75cm) high x 2-3/4" (6.99cm) deep; **W41 Backbox:** 4-11/16" (11.91cm) square electrical box with one-gang adapter for a total depth of 2-7/8" (7.30cm), supplied by others.

## ARCHITECTS' AND ENGINEERS' SPECIFICATIONS

### MODEL HPNCBES CODE BLUE EMERGENCY STATION

The wall recessed or surface mounted code blue station shall be a blue combination call and reset button with indicator lamp for originating emergency calls. It shall be used by hospital staff members to summon help in emergencies (primarily in Cardiac Arrest situations). Emergency call shall be cancelable only at the code blue station where it originated.

A rocker type button shall be used to operate a positive action toggle type switch which shall be provided with heavy duty independent contacts.

The mounting bracket shall permit correction of minor misalignment of backbox in wall. The high impact molded thermo-plastic faceplate shall have non-shattering and low water absorption characteristics and shall be self extinguishing rated 94V-0 by UL. Nylon screws shall be provided to mount the faceplate for maximum safety against electrical shock.

A self-adhesive "STAFF USE ONLY" metalcal, Part No. A44189, shall be available separately to be affixed to the faceplate where required.

The faceplate shall be finished in off-white. The S41 housing shall be a one-gang metal surface housing finished in alumex. The maximum dimensions shall be: **faceplate** - 2-3/4" (6.99cm) wide x 4-1/2" (11.43cm) high; **S41 housing** - 2-7/8" (7.30cm) wide x 4-5/8" (11.75cm) high x 2-3/4" (6.99cm) deep; **W41 backbox** - 4-11/16" (11.91cm) square electrical box with one-gang adapter for total depth of 2-7/8" (7.30cm), supplied by others.