



**AUTH-FLORENCE ELECTRONICS**

Designers and Manufacturers of  
Security • Mail • Communications Systems  
100 S. Main Street • Area A • Crystal Lake, IL 60014  
Telephone 815-356-3100 • 815-455-3783 • Fax 815-455-3820

# NC-9100

## Tone-Visual Nurse Call System

**UL 1069 LISTED**

**DOME LIGHTS  
NL-1/NL-2/NL-3/NI-2**

**NL-1**

**NL-2**

**NL-3**

**NI-2**

Auth-Florence NL and NI series Dome Lights are used in the NC-9100 Nurse Call System to provide visual indications for calls in the system. Corridor lights may be utilized with patient stations and with stand-alone emergency and code blue stations.

### Operation

When a patient station is activated, a steady white light indicates a normal call. A flashing white or red light indicates an emergency call. A rapidly pulsating red light indicates a code blue or smoke detector call.

### Equipment

The dome light is equipped with a wedge shaped translucent white plastic dome that snaps into a 2 gang plastic face plate for flush mounting. This plate mounts to a subplate, containing one, two, or three removable lamps corresponding to Models NL-1, NL-2, and NL-3 respectively.

The Model NI-2 Dome Light with Relay shall contain white and red colored lamps with an interface relay for use with a smoke detector. The contacts of the smoke detector turn on the relay inside the corridor light which originates a code blue to annunciator(s) and causes the red lamp in the station to rapidly pulsate on and off.

### Specifications

#### Front Plate:

Plastic 2 gang plate which mounts to back plate.

#### Back Plate:

Subplate construction that mounts either to single or two-gang electrical back box. Minimum back box depth shall be 2".

#### Lamp(s):

One, two, or three lamp models are available, NL-1, NL-2, NL-3 respectively. Lamps are replaceable from the front by removing the dome cover. Lamp's shall be #1829 bayonet base, 24 Volts. Lamp covers are used for red, blue, amber, and green colors (specify).

#### Dome:

Shatterproof, wedge shaped, translucent white plastic dome that snaps into the plastic face plate. Segment dividers in domes shall be provided for distinct lamp color separation.

#### Terminations:

Color coded wire leads.

